

Campbell Associates Ltd

5b Chelmsford Road Industrial Estate
GREAT DUNMOW, Essex, GB-CM6 1HD

www.campbell-associates.co.uk

Phone 01371 871030 Facsimile 01371879106



Certificate of Calibration and Conformance

CALIBRATION

0789

Certificate Number:- U18664

Test object: Sound Level Meter, BS EN IEC 61672-1:2003 Class 1 (Precision) and associated Frequency Analyser BS EN IEC 61260, Class 1

Manufacturer: Norsonic

Type: 131

Serial no: 1312731

Customer: Campbell Associates Ltd

Department: Hire

Address: Sonitus House,
5B Chelmsford Road Industrial Estate,
Great Dunmow, Essex. CM6 1HD.

Contact Person: Mr. David Egan - Laboratory Manager

Method :

Calibration has been performed as set out in CA Technical Procedures TP01 & 02 as appropriate. These are based on the procedures for periodic verification of sound level meters as set out in BS EN IEC 61672-3:2006 and for electrical testing of frequency filters as set out in BS EN IEC 61260. Results and conformance statement are overleaf and detailed results are in the attached Test Report.

| | Producer: | Type: | Serial No: | Certificate number |
|--------------|-----------|-------|------------|--------------------|
| Microphone | GRAS | 40AE | 86534 | 18663 |
| Calibrator* | Norsonic | 1251 | 30873 | U16836 |
| Preamplifier | Norsonic | 1207 | 12259 | Included |

Additional items that also have been submitted for verification

| | | |
|-----------------|----------|---------|
| Wind shield | Norsonic | Nor1451 |
| Attenuator | None | |
| Extension cable | None | |

These items have been taken into account wherever appropriate.

| | | | |
|---------------------------|-------------|--------------|--------------------|
| Environmental conditions: | Pressure: | Temperature: | Relative humidity: |
| Reference conditions: | 101.325 kPa | 23.0 °C | 50 %RH |
| Measurement conditions: | 100.40 kPa | 20.5 °C | 40.0 %RH |

Date received : 29/04/2015
Date of calibration: 29/04/2015
Date of issue: 30/04/2015

Engineer


Palanivel Marappan B.Eng (Hons), M.Sc

Supervisor


Darren Batten Tech IOA

This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to recognized national standards, and to the units of measurement realized at the National Physical Laboratory or other recognized national standards laboratories. This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.

Certificate of Calibration and Conformance

UKAS Laboratory Number 0789

Certificate Number:- U18664

Conformance

From markings on the sound level meter or by reference to the manufacturer's published literature it has been determined that the instrument submitted for verification was originally manufactured to BS EN IEC 61672-1:2002 and similarly that the associated sound calibrator conforms to BS EN IEC 60942.

Statement of conformance

The sound level meter submitted for testing has successfully completed the class 1 periodic tests of BS EN IEC 61672-3:2006, for the environmental conditions under which the tests were performed. As public evidence was available¹, from an independent testing organisation responsible for approving the results of pattern evaluation tests performed in accordance with BS EN IEC 61672-2:2003, to demonstrate that the model of sound level meter fully conformed to the requirements in BS EN IEC 61672-1:2002, and that the sound level meter submitted for testing conforms to the class 1 requirements of BS EN IEC 61672-1:2003.

The filter functions have been found to conform, by electrical testing, to the relative attenuation requirement of BS EN IEC 61260 (as required by UKAS Lab23) for a class 1 filter over the range of frequencies shown in the attached test report.

¹ This evidence is held on file at the calibration laboratory

Measurement Results:

| | |
|--|--------|
| Indication at the calibration check frequency - IEC61672-3 Ed.1 #9 | Passed |
| Self-generated noise - IEC 61672-3 #10 | Passed |
| Acoustical test of a frequency weighting - IEC 61672-3 Ed.1 #11 | Passed |
| Frequency weightings: A Network - IEC 61672-3 Ed.1 #12 | Passed |
| Frequency weightings: C Network - IEC 61672-3 Ed.1 #12 | Passed |
| Frequency weightings: Z Network - IEC 61672-3 Ed.1 #12 | Passed |
| Frequency and time weightings at 1 kHz IEC61672-3 Ed.1 #13 | Passed |
| Level linearity on the reference level range - IEC 61672-3 #14 | Passed |
| Toneburst response - IEC 61672-3 Ed.1 #16 | Passed |
| Peak C sound level - IEC 61672-3 Ed.1 #17 | Passed |
| Overload indication - IEC 61672-3 Ed. 1 #18 | Passed |
| Filter Test 1/1octave: Relative attenuation - IEC 61260, #4.4 & #5.3 | Passed |
| Filter Test 1/3octave: Relative attenuation - IEC 61260, #4.4 & #5.3 | Passed |
| Combined electrical and acoustical test - IEC 61672-3 Ed.1 #12 | Passed |

Comment

Correct level with associated calibrator is 113.9dB(A).

Observations

No information on the uncertainty of measurement, required by 11.7 of BS EN IEC 61672-3:2006 of the adjustment data given in the instruction manual or obtained from the manufacturer of the microphone was published in the instruction manual or made available by the manufacturer or supplier. The uncertainty of measurement of the adjustment data has therefore been assumed to be numerically zero for the purposes of this periodic test. If these uncertainties are not actually zero, there is a possibility that the frequency response of the sound level meter may not conform to the requirements of BS EN IEC 61672-1:2003.

The details of the uncertainty for each measurement is available from the Calibration Laboratory on request and is based on the standard uncertainty multiplied by a coverage factor K=2, providing a level of confidence of approximately 95%. The uncertainty evaluation has been carried out in accordance with UKAS requirements. Details on the sources of corrections and their associated uncertainties that relate to this verification are contained the detailed test report accompanying this certificate.

Calibration Report

Certificate Number:-18663

Manufacturer: GRAS
Type: 40AE
Serial no: 86534

Customer: Campbell Associates Ltd
Department: Hire
Address: Sonitus House,
5B Chelmsford Road Industrial Estate,
Great Dunmow, Essex. CM6 1HD.
Contact Person: Mr. David Egan - Laboratory Manager

Measurement Results:

| | Sensitivity: (dB re 1V/Pa) | Capacitance: (pF) |
|--------------------------|-------------------------------|----------------------|
| 1: | -25.57 | 13.8 |
| 2: | -25.56 | 13.8 |
| 3: | -25.56 | 13.8 |
| Result (Average): | -25.56 | 13.8 |
| Expanded Uncertainty: | 0.10 | 2.00 |
| Degree of Freedom: | >100 | >100 |
| Coverage Factor: | 2.00 | 2.00 |

The following correction factors have been applied during the measurement:
Pressure:-0.010 dB/kPa Temperature:-0.070 dB/°C Relative humidity:0.001 dB/%RH

Reference Calibrator: WSC1 - Nor1253-24269 Volume correction: 0.000 dB
Records:K:\C A\Calibration\Nor-1504\Nor-1017 MicCal\2015\GRAS40AE_86534_M1.nmf
Measurement procedure: TP05
All results quoted are directly traceable to National Physical Laboratory, London

The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor $k = 2$, which for a normal distribution corresponds to coverage probability of approximately 95%. The standard uncertainty of measurement has been determined in accordance with EA publication EA-4/02.


Comment:

Environmental conditions:

| | | |
|---------------------|---------------|--------------------|
| Pressure: | Temperature: | Relative humidity: |
| 100.489 ± 0.045 kPa | 21.7 ± 0.4 °C | 43.7 ± 2.6 %RH |


Date of calibration: 29/04/2015
Date of issue: 29/04/2015

Supervisor : Darren Batten TechIOA
Engineer :


Palanivel Marappan B.Eng(Hons), M.Sc
Software version: 6.0h


Campbell Associates
www.campbell-associates.co.uk

Microphone Calibration Certificate



GRAS
Type: 40AE

Serial no: 86534

Sensitivity: 52.70 mV/Pa
-25.56 \pm 0.10 dB re. 1 V/Pa
Capacitance: 13.8 \pm 2.0 pF
Date: 29/04/2015


Signature: *M. Panivel*

Measurement conditions:
Polarisation voltage: 0.0 V
Pressure: 100.49 \pm 0.05 kPa
Temperature: 21.7 \pm 0.4 $^{\circ}$ C
Relative humidity: 43.7 \pm 2.6 %RH
Results are normalised to the reference conditions.

Free field response
Diffuse field response
Pressure (Actuator) response

Campbell Associates
www.campbell-associates.co.uk

Microphone Calibration Certificate



GRAS
Type: 40AE

Serial no: 86534

Sensitivity: 52.70 mV/Pa
-25.56 \pm 0.10 dB re. 1 V/Pa
Capacitance: 13.8 \pm 2.0 pF
Date: 29/04/2015

Signature: *M. Panivel*

Measurement conditions:
Polarisation voltage: 0.0 V
Pressure: 100.49 \pm 0.05 kPa
Temperature: 21.7 \pm 0.4 $^{\circ}$ C
Relative humidity: 43.7 \pm 2.6 %RH
Results are normalised to the reference conditions.

Free field response
Diffuse field response
Pressure (Actuator) response

Campbell Associates
www.campbell-associates.co.uk

Comment: