

CONSTAT DE VERIFICATION VERIFICATION CERTIFICATE

N° CV-DTE-T-07-PVE-33833

DELIVRE A :
ISSUED FOR : RPS MANCHESTER

INSTRUMENT VERIFIE CHECKING INSTRUMENT

Désignation : Sonomètre Intégrateur
Designation : Integrator Sound Level Meter

Constructeur :
Manufacturer : 01dB-Metravib

Type : BLUE SOLO 01 N° de serie : 60843
Type : Serial number :

N° d'identification :
Identification number

Date d'émission : 12/12/07

Ce constat comprend 6 pages
This certificate includes pages

LE RESPONSABLE METROLOGIQUE
DU LABORATOIRE
THE METROLOGICAL HEAD OF THE LABORATORY

Philippe POURTAU



LA REPRODUCTION DE CE CONSTAT N'EST AUTORISEE QUE SOUS
LA FORME DE FAC-SIMILE PHOTOGRAPHIQUE INTEGRAL

THIS CHECKING REPORT MAY NOT BE REPRODUCED OTHER THAN IN
FULL BY PHOTOGRAPHIC PROCESS

CE DOCUMENT NE PEUT PAS ETRE UTILISE EN LIEU ET PLACE
D'UN CERTIFICAT D'ETALONNAGE. CE DOCUMENT EST
REALISE SUIVANT LES RECOMMANDATIONS DU FASCICULE DE
DOCUMENTATION X 07-011

THIS DOCUMENT CAN'T BE USED AS CALIBRATION
CERTIFICATE. THIS DOCUMENT IS MADE WITH
STANDARD X 07-011 RECOMANDATION.

CERTIFICATE OF CALIBRATION

ISSUED BY AV CALIBRATION LTD

Date of issue 11 December 2007 Certificate N° 03524



AV Calibration Ltd
13c Old Bridge Way
Shefford
Bedfordshire SG17 5HQ
England
Telephone (01462) 638600
Fax (01462) 638601
E-mail lab@avcalib.co.uk

Page 1 of 8 pages

Approved Signatory
G. Parry

A handwritten signature in black ink, appearing to be 'G. Parry', written over a horizontal line.

CLIENT RPS Group plc
52 Princess Street
Manchester
M1 6JX

F.A.O. Craig Williams

REF. - Job N° UKAS07/12334/01

DATE OF RECEIPT 30 November 2007

PROCEDURE AV Calibration Engineer's Handbook, Section 3: verification of sound level meters to BS 7580:Part 1:1997

IDENTIFICATION Sound level meter Rion type NL-31 [serial no. 01141927] connected via an extension lead type EC04A and preamplifier type NH-21 [serial no. 13572] to a half-inch microphone type UC-53A [serial no. 307684] fitted with a foam windshield type WS-10. Associated calibrator 01dB type Cal21 [serial no. 51231440(2003)] with one-inch housing and adapter type BAC21 for half-inch microphone.

CALIBRATED ON 10 December 2007

This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to recognised national standards, and to units of measurement realised at the National Physical Laboratory or other recognised national standards laboratories.
This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.